AGGREGATE

Grid Sieve Sets (Flakiness Index Sieve EN)



Used for the determination of the flakiness index of the aggregate. Consisting of an electrostatically painted frame and 5 mm diameter stainless steel bars.

PAG-410.1 Bar grid sieve opening 2,50 mm.

PAG-410.2 Bar grid sieve opening 3,15 mm.

PAG-410.3 Bar grid sieve opening 4,00 mm.

PAG-410.4 Bar grid sieve opening 5,00 mm.

PAG-410.5 Bar grid sieve opening 6,30 mm.

PAG-410.6 Bar grid sieve opening 8,00 mm.

PAG-410.7 Bar grid sieve opening 10,00 mm.

PAG-410.8 Bar grid sieve opening 12,50 mm.

PAG-410.9 Bar grid sieve opening 16,00 mm.

PAG-410.10 Bar grid sieve opening 20,00 mm.

PAG-410.11 Bar grid sieve opening 25,00 mm.

PAG-410.12 Bar grid sieve opening 31.50 mm.

PAG-410.13 Bar grid sieve opening 40.00 mm.

PAG-410.14 Bar grid sieve opening 50.00 mm.

Related Standards

EN 933-3, NF P18-561, UNI 8520-18, NLT 354

You May Also Be Interested In



Large Capacity Sample Splitter

PAG-412



Sample Splitter (Riffle Box)

PAG-418



Specific Gravity Frame

PGE-401

Review Product on Our Website

To access up-to-date technical details, images, and documentation for this product, you can scan the QR code on the right using your smartphone camera or visit our website directly.